

GST :06CCVPS2254R1ZJ



An ISO 9001: 2015 Co.

Aarson Scientific Works

9-10 Khojki Pur, P.O. Industrial Estate
Ambala Cantt. 133006 (Haryana) India

PH : 91-0171-2699426, MB : 09466214426, 9996736100

www.aarsonindia.com

Email : aarsons1980@gmail.com, aarson1980@gmail.com

aarsonexports@gmail.com

CHLORINE Comperator
(Aarson Make)
(Model : RSE-490)

Technical Data sheet



S.No.	DESCRIPTION	SPECIFICATION
1	Name of Equipment	Chlorine Comparator
2	Residual Chlorine Range	0.1 to 3 PPM
3	pH Range	6 to 8.4
4	Indicator	phenol Red Indicator Disk
5	Make	AARSON

For AARSON SCIENTIFIC WORKS

R. Singh
Prop